

12

EUROPEAN PATENT APPLICATION

21 Application number: 89301280.7

51 Int. Cl.4: G01N 27/26

22 Date of filing: 10.02.89

30 Priority: 16.02.88 US 156430

43 Date of publication of application:
 23.08.89 Bulletin 89/34

64 Designated Contracting States:
 AT BE CH DE FR GB IT LI LU NL


71 Applicant: APPLIED BIOSYSTEMS, INC.
 850 Lincoln Centre Drive
 Foster City California 94404(US)

72 Inventor: Lauer, Henk H.
 2707 Ponce Avenue
 Belmont California 94002(US)
 Inventor: Grossman, Paul D.
 590 Avocet Drive, No. 720
 Redway City California 97065(US)
 Inventor: Mead, Dennis E.
 1495 McBain Avenue
 Campbell California 95008(US)

74 Representative: West, Alan Harry et al
 R.G.C. Jenkins & Co. 26 Caxton Street
 London SW1H 0RJ(GB)

54 Capillary electrophoresis.

57 Capillary electrophoresis apparatus includes an electronically controlled valve system 97 for automatically introducing a sample 93 into the capillary 13 by means of a vacuum at the end of the capillary tube. Such sucking in of the sample is extremely accurate and reproducible, results in a minimum of band broadening and enables the entire system to be easily automated. Automated temperature control enables the temperature of the capillary tube (and hence the solvent/solute system) to be controlled during electrophoresis, thereby controlling pH and electrophoretic mobility. The capillary may be prewashed and equilibrated to achieve substantially zero charge on the capillary wall, thereby essentially eliminating electroosmotic flow and substantially improving resolution.



CAPILLARY ELECTROPHORESIS

This invention relates to capillary electrophoresis, or as it is more conventionally called "capillary zone electrophoresis" (CZE), and more particularly to automated methods and apparatus for introducing samples into capillary columns and for improving separations by using temperature control of said columns.

In recent years significant advances have been made in micro-column separation techniques. A principal advantage of such techniques is their suitability for analysis of extremely small sample volumes, eg. in the microliter or submicroliter amounts of sample. Being able to analyze such small volumes has become exceedingly important with the explosion of research in the biological field, because often-times biological samples are quite small.

One of the significant problems with capillary techniques is in introducing sample into the capillary. One technique used in capillary electrophoresis, called sample injection, is electromigration, a term collectively including the effects of electrophoresis and electro-osmosis (See Jorgenson, J.W. and Lukacs, K.D., *J. Chromatography*, 1981, Vol. 218, pp. 209-216; Jorgenson, J.W., and Lukacs, K.D., *Science*, 1983, Vol. 222, pp. 266-272; and Wallingford, R.A. and Ewing, A.G., *Anal. Chem.*, 1987, Vol. 59, pp.681-684). In this technique, one end of the capillary and the electrophoresis anode are placed into the sample and a voltage is briefly applied, causing a small band of sample to electromigrate into the capillary. This method of sample injection suffers from discrimination within the sample because solutes with higher mobilities will preferentially migrate into the electrophoresis column, and therefore change the relative composition of the sample. To avoid this problem, attempts to physically inject sample have also been reported (Jorgenson and Lukacs, *Science*, *ibid*). However these direct injection techniques cause band broadening, apparently due to the laminar flow profile introduced during the injection.

Other less common injection methods include gravity flow (See Tsuda, A., et al, *J. Chromatography*, 1983, Vol. 264, pp. 385-392.), siphoning (See Honda, S. et al, *J. Chromatography*, 1987, Vol. 404, pp. 313-320.), and the use of an electronic sample splitter (See Deml, M. et al, *J. Chromatography*, 1985, Vol. 320, pp. 159-165.). Each of these injection techniques are capable of placing subnanoliter volumes of sample into the electrophoresis column with a minimum of band broadening. However, the gravity flow or siphoning injection method is inaccurate and lacks precision in providing absolute volume amounts due to the unreliable position of the sample lever which will change due to the sample withdrawal. The latter can only be neglected if the original sample volume is large compared to the volume injected. With the electronic splitter, a larger initial sample volume is required in order to be able to split it down to the smaller size required for the column. Thus, some sample may be wasted, or there may not be sufficient sample to perform a separation. Also, this latter technique is further complicated by requiring an additional controlled power supply or very careful control of the electric resistances in the different legs of the splitter. Furthermore, the need to use an initial larger sample size significantly decreases the number of applications for which it can be used.

What is needed is a simple, automatable, sample injection technique that is suitable for microvolumes, is capable of providing accurate sample volumes, and which produces a minimum of band broadening.

In accordance with preferred embodiments of the invention, an apparatus is disclosed for providing capillary electrophoresis, which includes an electronically controlled valve system for automatically introducing a sample into the capillary by means of a vacuum at the end of the capillary tube. This approach of sucking in the sample is extremely accurate and reproducible, and results in a minimum of band broadening. Furthermore, it enables the entire capillary electrophoresis system to be easily automated.

The apparatus includes first and second reservoirs that are electrically isolated from each other for holding electrophoretic media, a sample reservoir located in proximity of the first reservoir for holding a sample to be electrophoresed, and a high voltage power supply connected between the first reservoir and the second reservoir. A first pressure source of a first gas having a first known pressure, typically the ambient air pressure, is used for providing an environment for the first reservoir and the sample reservoir, so that electrophoretic media in the first reservoir and sample in the sample reservoir are under the first pressure. The apparatus includes a pressure reservoir for holding a second gas (also typically air) having a second pressure that is lower than the first pressure. A capillary tube is also included in which to electrophorese the sample. A rack system is provided for holding the first and second reservoirs, the pressure reservoir, the high voltage power supply, the sample reservoir, and for holding one end of the capillary tube in the second reservoir. A gas connecting system connects the second reservoir to the pressure reservoir, the connecting system having a valve for venting the connecting system to the first pressure source and for blocking communication of the second reservoir with the pressure reservoir while venting to the first pressure source. The apparatus also has an insertion element for inserting the other end

of the capillary tube into the sample reservoir and into the first reservoir. In the preferred mode, the apparatus includes a computer system for controlling the insertion element and the valve, so that when the other end of the capillary tube is in the sample reservoir, the valve permits communication of the second reservoir with the pressure reservoir for a controlled period of time for sucking the sample into the capillary tube. Additionally, in the preferred mode, the computer system causes the other end of the capillary tube to be transferred to the first reservoir after the sucking of the sample into the capillary tube. After the sample has been introduced into the capillary tube and the tube has been transferred to the first reservoir, the electrophoresis is begun.

An additional important feature of the preferred embodiment is that an automated temperature control system is provided which enables the temperature of the capillary tube (and hence the solvent/solute system) to be controlled during electrophoresis. This is particularly advantageous in that for many buffers, the pH is a strong function of temperature; hence temperature control is very directly pH control. Additionally the pH can have a direct effect on the electrophoretic mobility and hence the separation efficiency.

In another embodiment of the invention, the capillary is prewashed and equilibrated to achieve substantially zero charge on the capillary wall, thereby essentially eliminating electroosmotic flow and substantially improving resolution.

Fig. 1 shows an apparatus according to the invention.

Fig. 2 is a table that illustrates the effects of capillary prewash on electroosmotic mobility.

Fig. 3 is a graph showing the results of Fig. 2.

Fig. 1 is a partially sectioned illustration of a preferred embodiment of an automated capillary electrophoresis, henceforth CZE, apparatus according to the invention. In this preferred mode, the apparatus includes an environmental enclosure 11, which has access openings (not shown), and feedthroughs of various kinds through the walls of the enclosure for elements that must be connected to elements outside the enclosure. Electrophoresis is accomplished within the enclosure in a capillary tube 13, preferably constructed of fused silica, such as is typically used for high sensitivity liquid or gas chromatography. One end of capillary 13 is immersed for the process in a buffer solution 19 held in a first container 21, and the other end is immersed in a buffer solution 15 in a second container 17. Buffer solutions 15 and 19 are typically the same solution, and many are well known in the art. Although in the prior art many different pH's have been used for the various buffers, depending on the particular experiment being performed, in this preferred embodiment, it has been found that for capillary electrophoresis that a relatively low pH is best. In the preferred mode, to achieve the best separations, the pH is adjusted to the point of zero electric charge of the buffer-capillary combination, ie. the point at which there is no charge on the capillary wall. As will be discussed subsequently in more detail, the point of zero charge will vary depending on the buffer used and the pretreatment of the column. However, as a practical matter, typically a pH below about 2.5 will suffice. Also, as will be discussed subsequently, sometimes particular buffers are used which are temperature dependent, ie. their pH varies strongly with temperature, or stated another way dpK/dT is relatively large.

View 23 is an enlargement of capillary tube 13 in cross section. The internal diameter of the capillary, D_1 , varies for different kinds of samples and for other reasons. A typical value for D_1 is 0.05 mm, and generally varies between zero and 200 microns. The wall thickness of tube 13 is small enough that the tube is flexible and may generally be manipulated without breaking. Also, the small diameter allows for efficient heat transfer.

Second container 17 has an airtight top 25. Capillary tube 13 enters the second container through a stopper 27 maintaining an airtight seal. There are two additional penetrations through top 25. A hollow tubing 29 enters through stopper 31 and an electrode 33 enters through another stopper 35. Stopper 35 is typically of an electrically non-conducting material. From electrode 33, an electrical lead 37 goes to an electrical feedthrough 39 which allows an electrical signal or power to cross the wall of the enclosure without shorting to the enclosure. On the outside, electrical lead 41 goes to a terminal of a high voltage power supply 43.

From the opposite terminal of power supply 43 another electrical lead 45 goes to another feedthrough 47. Inside the enclosure lead 49 goes to an electrode 51 immersed in buffer solution 19 in first container 21. With buffer solution and sample material in the capillary tube and the tube ends immersed in buffer solution in the two containers, power supply 43, through the electrical leads, feedthroughs and electrodes, may be used to maintain an electrical potential across the material in the capillary tube.

The second container rests on a support 53 with an electrical insulator 55 between the container and the support. The insulator is needed if the container and support are electrically conductive. First container

21 rests on a moveable, sliding support 57 and an insulator 59.

A detector 61 is positioned relative to the capillary tube to measure the results of electrophoresis in the capillary. Such detectors are well known in the art, and include for example an Applied Biosystems Model 783 Spectroflow UV/Visible Detector, which is a variable wavelength programmable detector that is specifically adapted for on-column detection. Electrical leads through feedthroughs 63 carry power and signals for the instrument. There may be more than the two leads shown.

When the electrophoresis process is complete on one sample, and another sample is wanted in the capillary for analysis, a new sample may be loaded without manual intervention or disturbing the environmental enclosure. A motor 65 powered by leads through feedthrough 67 and supported by bracket member 69 may be activated to turn lead screw 71. Nut 73 is attached to member 75 with a clamp 77 securely holding tube 13, so that turning lead screw 71 will raise and lower the tube by the distance between stops 79 and 81. This distance is set to be sufficient for the lower end of capillary tube 13 to be raised above the rim of container 21, and lowered again.

With tube 13 raised above the rim of container 21, motor 83 may be activated by leads through feedthroughs 85 to turn lead screw 87 moving slide 57 along support 89. A sample container 91 with multiple microvolumes such as 93, arranged in a row in the container, is prepared in advance and placed adjacent to container 21 on slide 57. Each microvolume may contain a sample to be analyzed. Typical injection volumes range from 1 nl to 10 nl in this preferred mode, although other size samples could of course be chosen depending on the size of the reservoir used to hold the sample and the size of the column. By controlling motor 83 moving slide 57, any one of the microvolumes of container 91 may be moved to be directly below the end of capillary tube 13, which may then be lowered into the microvolume by control of motor 65. Once a new sample is drawn into the capillary, the capillary may again be raised, container 21 returned to position, and the end of the capillary re-immersed in the buffer by lowering the tube.

To inject a new sample, while one end of the capillary is in one of the microvolumes of sample material, a relative vacuum is drawn in second container 17 by means of tubing 29 which exits the environmental enclosure. Motor 95 is controlled to rotate a three-way rotary valve 97 opening tubing 29 to vacuum reservoir 99. The reservoir is maintained at desired vacuum level by vacuum pump 101 through isolation valve 103. A vacuum sensing gauge 115 with programmable signal points monitors the vacuum level in reservoir 99. The pump is powered by motor 105. Careful control of timing and vacuum level provides a very accurate method for drawing a predetermined amount of sample material into the capillary, as well as other benefits. As an example, using a pressure differential of 5.0 in. of Hg between the vacuum reservoir and the enclosure 11, with a 65 cm long fused silica capillary having a 50 micron inside diameter, a 2 second open time for valve 97 results in an injection quantity of 5 nanoliters of an aqueous solution.


Another important feature of the apparatus according to the invention is that the temperature inside the environmental enclosure 11 can be controlled. A heating element 107 is powered through feedthroughs 109 to provide heat, and a heat sensing element 111 monitors temperature through leads 113. As will be discussed subsequently, such provision for temperature control is very useful, since some buffers have a temperature dependent pH, and for those buffers pH can be controlled automatically by controlling temperature. Temperature control is also useful in the more general case since other kinds of variations are avoided if a uniform temperature is used throughout a separation. For example, viscosity and therefore mobility are most often strong functions of temperature, so that for reproducibility, temperature control is required.

Power and control leads for all the electrical equipment associated with the apparatus of the preferred embodiment are carried by electrical conduit 121 to a control interface 119 which provides power terminations and switching of signals for control purposes. The control interface is connected to and manipulated by a computer 117 which can be pre-programmed so that critical parameters may be maintained and sequences of analyses may be performed automatically by the apparatus. For example, the vacuum level desired can be entered as control data, and the computer, through the control interface, monitors the signal from vacuum gauge 115 and opens and closes vacuum isolation valve 103 so that the desired vacuum level is closely maintained. As another example, the computer can be used to control the temperature inside the environmental enclosure by monitoring temperature sensor 111 and controlling power to heating element 107 as needed to maintain the programmed temperature. Also, the computer can be programmed to allow a sequence of analyses to be made, using the several samples preloaded into microvolumes in container 91, controlling the electrical devices in the required sequence. The program may be set to run analyses on all of the microvolume samples, one-after-the other, or to allow for manual intervention and initiation between each analysis. Another important feature of computer control is that the computer makes it possible to reverse polarity of the capillary electronically. Hence, for solutes that are of

opposite charge, one can reverse the direction of migration of solute particles and thereby use the UV detector at its fixed location. Details of the program structure for the computer are provided in Appendix A.

5 Control of pH

In capillary electrophoresis in free solution and in some gels, solutes with different charges (absolute) have different electrophoretic mobilities and can therefore be separated. Separation efficiency can be improved if the selectivity (ie. the relative difference in electrophoretic mobility) between two or more solutes can be changed during the electrophoretic run. One way to achieve this is to change the relative difference in the effective charges on the species to be separated. In many cases the pH (or better, the pK) of the solution in the capillary, which is mostly buffer, determines the charge on solutes that obey the rules of acid/base equilibria. For example, for CHES (Cyclohexyl amino ethane sulfonic acid) in water, a chemical equilibrium is established which is dependent on the particular temperature, as indicated by the following formula:



at pH=9.55, 50% of the CHES is in the zwitterion form and 50% is in the anion form. By increasing the pH to 10.55, the anion (RSO_3^-) concentration will be ten times that of the zwitterion and by increasing it to 11.55, the equilibrium will be pushed almost completely to the anion side. At that point only 1% of the CHES will exist as a zwitterion. The anion will possess a certain electrophoretic mobility while the zwitterion being electrically neutral will not have an electrophoretic mobility. What this means is that at a pH of about 7.55, the CHES solute will have practically no electrophoretic mobility and at a pH of about 11.55, it will have nearly the mobility of the anion. Hence, by changing the pH of the solution, the mobility of a solute can be changed.

As indicated earlier, in many cases, the pH (or pK) of buffer solutions are a function of their temperature, and different buffer solutions have different temperature characteristics. (See CALBIOCHEM CATALOG, Table IV, Page 16.) By changing the temperature of the buffer in time or in space, different pH's in time and in space can be generated and thus the mobility of a species can be manipulated. In general, the pK of a solution is given by:

$$\text{pK} = \text{pH} - \log\{\text{RSO}_3^-/\{\text{R}^+\text{SO}_3^-\},$$

where $\{\text{RSO}_3^-\}$ corresponds to the concentration of the anion RSO_3^- , etc., and in many cases the pK has a strong temperature dependence. As an example of how to use this temperature dependence during the performance of a separation, suppose that a separation is to be performed on a solution containing three solutes; A, B, and C, and that the A and B separation is best performed at a first temperature T1 with A coming through first, and that the B and C separation is best performed at a temperature T2. The separation is run for a first time at temperature T1, until A is separated from B and C, then the temperature is changed to T2 until B and C are separated. Similarly, more complicated temperature profiles can be used depending on the particular solutes being separated, for example continuous programming can be used if desired.

It should be appreciated by those skilled in the art that solutes that are to be separated may also obey the acid/base equilibria rules, and as a result also can change their degree of ionization with temperature. This solute effect will be superimposed on the pH change of the solvent (buffer) and hence, depending on the choice of buffer and solute combination, can provide an enhanced mobility difference, decreased mobility difference, or no change in mobility difference at all. Hence, various combinations of buffer and solute should be chosen to achieve the desired effect.

As a specific example of the effects of varying temperature, and hence pH, an experiment was conducted to investigate the relative electrophoretic mobilities of two proteins, Myoglobin (wsm) and Myoglobin (hh). A fused silica capillary was used having a 55 cm length to the detector, a total length of 70 cm, and an inside diameter of 0.050 mm. Using a 10 mM Tris-HCl buffer, and a 20 kV electric potential, the change in relative difference in electrophoretic mobilities (ie. selectivity) of the two proteins was measured between the temperatures of 28.9 C (pH=8.90) and 62.4 C (pH=7.90), and was found to be minus 45%.

As described earlier, another important aspect of the invention in achieving a high selectivity, particularly in protein separations, is to eliminate charge on the capillary wall. The purpose is to eliminate electroosmotic flow, so that the column is not being swept during the run, thereby providing a longer separation time in the column (lower average velocity of the species to be separated), and hence better resolution. Also, by eliminating charge on the wall, positive ions (eg. proteins) do not stick to the wall, unlike the typical case when the wall is negatively charged. One way to achieve zero charge on the wall is through control of pH. Generally, the electroosmotic velocity is proportional to the zeta potential times the applied electric field divided by the viscosity. The zeta potential describes electrostatic forces in the interfacial double layer between two phases and is, among others, a function of the differential adsorption of ions. When there is no electroosmotic flow, the zeta potential is zero, and there is no charge on the capillary wall. Hence, by measuring the electroosmotic mobility, ie. the electroosmotic velocity divided by the applied field, as the pH is changed to achieve zero mobility, the point of zero charge on the wall can be determined.

An example of the effects on electroosmotic mobility resulting from pH changes and some surprising results from differences in capillary preparation are illustrated in the Table of Fig. 2, and in Fig. 3. In these experiments, a number of buffers were used in order to cover a wide range of pH levels, since as a general rule any one buffer has a relatively limited range of pH values over which it is useful. The capillary used was a fused silica capillary supplied by Polymicro Technologies, and had a length to the detector of 30 cm, a total length of 50 cm, and an inside diameter of 0.050 mm. The applied electric field used was 360 V/cm. Except for run 4, where the order of preparation was reversed to check for reversibility (ie. steps 4, 5, and 6, preceded steps 1, 2, and 3 below), the protocol for capillary preparation was as follows:

1. wash capillary with 1 M NaOH for 3 minutes;
2. equilibrate with buffer for 5 minutes;
3. run with mesityl oxide as a neutral marker and measure elution time for the marker;
4. wash capillary with 1 M HCl for 3 minutes;
5. equilibrate with buffer for 5 minutes;
6. run with mesityl oxide and measure elution time.

The buffers used were CAPS (ie. 3-(cyclohexylamino) propane sulfonic acid), BICINE (ie. N,N-bis (2-hydroxyethyl) glycine), MES (ie. a sodium salt of 2- (N-morpholino) ethane sulfonic acid), and citric acid.

A graph of the electroosmotic mobility, shown in Fig. 3, illustrates dramatically the results of lowering the pH. Clearly, as the pH is lowered there is a sharp decrease in mobility indicating that the charge on the wall is quickly approaching zero. For NaOH treated tubes, a pH of below about 2.5 corresponds essentially to zero electric charge on the wall (the baseline noise is proportional to the current, so as a practical matter it is important to use a low current, and low pH buffer). Even more striking, however, is the effect of the prewash. Although NaOH is the quintessential strong base that is typically used for washing glass, it is clear that the use of a strong acid such as HCl is a much better prewash if the purpose is to eliminate charge on the capillary wall. For example, if the capillary prewash is performed with HCl, a pH of about 4.0 eliminates about 97% of the charge on the capillary wall, and produces an electroosmotic mobility that is even lower than that achieved using a pH of 2.5 if the prewash is with NaOH. Furthermore, it appears that the effects of the different prewashes are substantially independent of each other, since in run 4 where the order of the two prewashes was reversed, the results are substantially the same. As a practical matter, it appears that removing 95% or more of the charge on the capillary wall is important in performing high resolution capillary electrophoresis, regardless of the prewash that is used. However, it appears that eliminating that charge is much easier to accomplish, and allows use of a much higher and more easily attained pH level, if the capillary is prewashed with an acid instead of a base before the run.

It will also be appreciated by those skilled in the art that there are several ways to control the temperature of the solvent/solute system. For example, one way has already been described which uses a heater system for environmental chamber 11. Another approach would be to use one or more electrical heaters wrapped around the capillary tube, and another would be to use one or more pieces of insulating wrap on the capillary tube. Those skilled in the art will undoubtedly be able to think of other equivalent methods for controlling the temperature to effect electrophoretic mobility. Those skilled in the art will also understand that in some instances it may be preferred to not have all components inside the enclosure 11. For example, the detector sometimes may be located outside the enclosure along with the corresponding portion of the capillary where the UV detection is to take place. Such an approach would facilitate service of the UV detector system. Also, instead of raising and lowering the capillary, one could raise and lower the sliding support to insert and remove the capillary from the sample and buffer reservoir. It should also be apparent that one could use electrophoretic media other than aqueous solutions, for example organic fluids

could also be used, a specific example being acetonitrile.


I. Operational Description:

5

The software will have three primary operating modes. These are Idle, Run, and Manual. Each of these modes relates to a basic state of operation for the instrument. Each of the primary modes have a series of associated sub-modes.

10

A. Functional Layout:



50

B. Mode Descriptions:

Each Mode is selected by using the special function keys to the right of the display. Choices are presented to the chemist on the display and by using the NEXT and PREVIOUS keys changes the appropriate values in the appropriate units for the particular choice being edited or selected. To change from choice to choice the Left and Right Arrow keys are used.

55

The following are brief descriptions of the instrument's modes.

1. Idle Mode:

This is the power on and default operating mode of the instrument. Any of the other modes may be started from idle, and return to idle when they are exited. Idle allows the user to execute the functions listed below.

- a) Cycle Edit
- b) Method Run
- c) Manual Control
- d) Configuration

Doors may be opened in Idle mode at anytime.

a. Edit Method:

Allows the chemist to display, and modify saved methods or to copy a method into a new method name for editing. The system will maintain in [battery backed up] RAM between 1 and 20 methods. The system will contain in ROM example executable methods [1-20] which can be copied into RAM and be edited. This function will allow the chemist to access a method, modify it and save it back. All method parameters will be validated at input, and prior to saving. The list of method parameters, and their meaning will be presented later.

Editing of any RAM based method can occur during a method run but will not change the method run in process. Changes made to the method in process will be seen the next time that method is run.

b. Run Method:

Allows the chemist to execute one of the saved methods automatically. Execution performs the following set of cycles, in the order shown. Doors can be opened only when a pause state is in process. With the combination of RAM and ROM methods there will be a total of 40 methods available to the chemist.

i)	Interlocks	All interlocks are verified.
ii)	Wash	Capillary washed w/ reagent.
iii)	Equil Buffer	Buffer is brought into the capillary and equilibrated.
iv)	Marker Inj.	Marker solution injected.
v)	Sample Inj.	Sample solution injected.
vi)	Detector	Initialize Detector
vii)	t0	Initial Time Step.
viii)	t1	First Change Option.
ix)	t2	Second Change Option.
x)	t3	Third Change Option.
xi)	tEnd Run	Run continues until this time expires, or stop/abort pressed.

c. Manual Control:

In this mode the chemist can set Instrument parameters from the key board with realtime results. Introduction of solutions from vials is allowed, but the sequence of solution introduction is carefully controlled when HIGH VOLTAGES are involved. High voltage is allowed only when capillary is in a sample or buffer position, and all safety interlocks are verified. Doors may be opened in the manual mode when no voltage is applied to the capillary and the vial holder is not moving.

d. Configuration:

This allows the chemist to set default values for the system parameters. This includes all detector parameters.

Self test is executed at power up doing only those tests that do not require user intervention, and whenever diagnostics are required all tests are available.

- 5 Adjustment of capillary is accomplished at this state allowing the vial holder to be in the up position with the doors open.

10 II. Methods:

A. Description of Parameters:

15 The ACE will be capable of running up to four (4) samples unattended by the chemist. To do this the system must know in advance the values for a set of control parameters. This set of parameters will be called an ACE method.

The following is a list of the parameters in an ACE method.

- 1) Wash Time
- 2) Buff Equil Time
- 20 3) Marker Inj Time
- 4) Sample Inj Time
- 5) Polarity
- 6) Wavelength
- 7) Range
- 25 8) Risetime
- 9) Autozero
- 10) Time
- 11) Buffer Number
- 12) Voltage
- 30 13) Temperature

Each of these parameters is used during an automated run to control the instrumental conditions. [The last four parameters are optionally changeable three times during an ACE method run.]

35 Some of the parameters are not currently set by the chemist, but use instead preset values determined by ABI/SC chemists.

- 1) Temperature Equil Time
- 2) Inject Vacuum Level
- 3) Inject Voltage

40

1. Wash Time:

45

Range:	0 - 60 min
Default:	5 min
Unit:	1 min

50 The wash time is how long a strong buffer is applied to the capillary as the first step in an automated run. If this time is set to zero, this step in the run sequence will not occur.

2. Buffer Equilibration Time:

55

EP 0 329 341 A2

Range:	0 - 20 min
Default:	1 min
Unit:	1 min

5

The third step of the run procedure is to introduce the buffer into the capillary. This is performed by a vacuum sucking on one end of the capillary. This time is variable, and can be set by the chemist. Temperature equilibration could take place at the same time possibly. If this time is set to zero, this step in the run sequence will not occur.

10

3. Marker Injection Time:

15

Range:	0 - 30 sec
Default:	1 sec
Unit:	25ms accuracy .5 second increments

20

The marker solution is introduced into the capillary at the end of the equilibration state. A very small amount (a few μL) of this solution is introduced as a plug into the end of the capillary, by using a lower vacuum for a controlled period of time.

25

4. Sample Injection Time:

30

Range:	0 - 30 sec
Default:	3 sec
Unit:	25ms accuracy .5 second increments

35

The sample solution is introduced next into the capillary as a second small plug. Again the volume injected is a function of vacuum level, time, viscosity and capillary dimensions.

5. Polarity:

40

Range:	A
Default:	Anode [at sampling side]
Unit:	N/A

45

The polarity of the sample side buffer vial can be either positive or negative. This value is set in the method and is maintained for the length of the run.

6. Wavelength:

50

Range:	190 - 700 nm
Default:	215 nm
Unit:	1 nm

55

The absorbance wavelength of the detector for the run. This value is user set at the start of the run, and

remain constant through out the run.

7. Range:

5

Range:	0.001 - 3.000 AUFS
Default:	0.01 AUFS
Unit:	0.001 - 0.1 is 0.001; 0.1 - 3.000 is 0.01 AUFS

10

The absorbance units full scale of the detector is user set at the start of the run and remains constant for the duration of the run.

15

8. Risetime:

20

Range:	0.02 - 5.0 secs
Default:	.5
Unit:	8 steps

The detector risetime (response time) is user set at the start of the run, and remains constant for the duration of the run. The value has 8 incremental values between 0.02 and 5.0 seconds.

25

9. Autozero:

30

Range:	Yes/No
Default:	Yes
Unit:	N/A

35

For the present system the detector will be autozeroed at a set time after the voltage is turned on in the run. [See T Zero sequence]

40

10. Time:

45

Range:	0.0 - 6.0 mins
Default:	25 mins
Unit:	1 min

The time along with the voltage, and temperature will be used to create a series of steps for the run to follow. This will be defined in the next section.

50

11. Buffer Number:

55

EP 0 329 341 A2

Range:	1 - 2
Default:	1

5 There are two (2) buffer positions on the vial tray. Each method will need to know which buffer(s) is to be used with the run. [Both buffers can be used in the same run.] When buffer change is selected during a run the software will select the opposite buffer automatically the user will not need to specify a buffer number. When a buffer change is selected the voltage will be turned off during the change and will be turned on.

10

12. Voltage:

15

Range:	0 - 30 KV
Default:	15 KV
Unit:	1 KV

20 The value of the voltage is the potential across the ends of the capillary at the time specified. This value may range from 0 to 30 kilovolts.

13. Temperature:

25

Range:	15 - 60 °C
Default:	30 °C
Unit:	1 °C

30

35 The value of the temperature is the temperature of the thermostated oven in the instrument at the time specified. It may range from 15 to 60 °C. If the value entered is below 25 °C than a caution will be displayed that coolant must be present to achieve desired temperature. In the future the software with a hardware ambient temperature detector will give this caution according to the temperature selected and the actual ambient temperature.

A.1 Fixed Parameters:

40

Some parameters are not currently set by the chemist, but use instead preset values determined by ABI/SC chemists.

1. Temperature Equilibration Time:

45

Future Range:	0 - 10 min
Fixed:	6 min

50

55 The primary step of an automated run is bringing the oven up to the runtime temperature. This value is set in the software, and cannot be changed by the chemist. If temperature is achieved prior to expiration of this time the run proceeds to the next step.

2. Inject Vacuum Level:

Range:	0 -20 inches Hg
Default:	5 inches Hg
Future Unit:	1 inch

5

The vacuum level which is used to introduce the sample, and marker into the capillary is set in the software, and copied into the method when it is first created.

10 3. Inject Voltage:

15

Range:	0 - 30 KV
Default:	6 KV
Future Unit:	1KV

20

The voltage level at which sample and marker is electrophoresed into the capillary is set in the software.

B. Method Execution:

25

Section A above describes the Instrumental parameters involved in an automated run performed on the ACE. This section will briefly describe cycle particular information and operation.

30

When the chemist selects the RUN method function, the software asks which method to run. A number between 1 and 20 is entered for the desired method. That method is then copied from the method buffers into the run buffer. At this time the chemist is asked the number of Samples to analyze using this method. The Instrument then begins execution of the run buffer.

1. Temperature Equilibration:

35

An optional Pause can be performed before this step continues. [See Pause Note]

40

The second state in a run sequence is temperature equilibration. Oven is to start equilibration as soon as value is entered [Manual Mode] or as soon as a method number is selected [Auto Mode]. The oven temperature is set from the value in the time zero (0.0) step of the method. A period of up to six minutes is allowed for the oven to reach this temperature. The temperature is determined to have been reached when: 10 [to be determined in actual use most likely a value of 2] sequentially acquired temperature values have delta between the set value, and the read value of $\pm .1^{\circ}\text{C}$ precision; $\pm .2^{\circ}\text{C}$ accuracy; read $\pm .25^{\circ}\text{C}$ for 10 seconds .

When this condition is met, the run advances to the next state.

45

2. Wash:

An optional Pause can be performed before this step continues. [See Pause Note]

50

The first state in a run sequence is the WASH state. This state draws a strong acid or base solution through the capillary. This is done by positioning the capillary in the wash vial. The capillary is then exposed to a vacuum at the wash fixed level. This exposure has a time duration as set in the wash time.

3. Buffer Equilibration:

55

An optional Pause can be performed before this step continues. [See Pause Note]

The third state of a run is the equilibration of the buffer to the capillary. This is accomplished by moving the capillary to the buffer designated in the method. The buffer is then drawn through the capillary at the designate wash vacuum level, for the designated buffer equilibration time.

4. Marker Injection:

An optional Pause can be performed before this step continues. [See Pause Note]

5 The fourth state of a run is the injection of the marker solution into the capillary. This is accomplished by positioning the capillary into the marker vial. The marker is then drawn into the capillary for the designated injection time. The method of injection is by exposing the capillary to a vacuum at the injection level, or a voltage at the injection level.

5. Sample Number:

Range:	1 - 4
Default:	1

15 There will be four (4) sample positions on the vial holder. At the time a method run is selected the number of Samples to be analyzed will also be selected.

6. Sample Injection:

An optional Pause can be performed before this step continues. [See Pause Note]

20 The fifth state of a run is the injection of the sample into the capillary. The capillary is positioned to the designated sample vial, and the processed the same as a marker, except the exposure time is that of the sample inject time.

7. Begin Final Run:

30 An optional Pause can be performed before this step continues. [See Pause Note]

At this state actual electrophoresis begins. The polarity of the buffer is set per the method and the voltage is set to the level specified in the time zero (0.0) step of the method. A specified number of seconds after the voltage is turned on, the detector is sent the command to autozero itself, and the remote integrator event is executed.

Each method can have up to four (4) gradient steps associated with it. These steps are contain the information shown below:

Time: Voltage: Temperature:

40 The method steps are executed in time sequential order. Every method has a time 0.0 step. This sets the initial voltage and temperature for the run. If no other steps are entered, these conditions are maintained for the length of the run. If other steps are present the voltage and/or temperature are stepped over the time period between sequential method steps. The time of the final step is the total runtime of the method. For instance the following method would have a runtime of 30 minutes, and would have two state changes [voltage & or temperature].

Time:	Voltage:	Temperature:
0.0 min	20 KV	30 ° C
10.0 empty	30	30
20.0	30	empty
		40
30.0	30	40

7. End of Run:

When the runtime has been reached, the voltage is turned off. At the same time the remote integrator event is executed again. The instrument then returns to the Idle mode. The temperature that was T zero temperature is maintained as the idle temperature before the next run.

5

8. Run Functions:

There are four functions available to the chemist during an automated run. The first of these is autozero. By executing this function, the chemist sends the detector the instruction to autozero itself. The second function when executed stops the methodmed run early. The third function aborts the methoded run at
10 anytime in the sequence. The fourth function allows the chemist to go from automated control to manual control at the current conditions of the instrument.

15 Pause Note:

SemiAutomatic operation will follow the same sequence as the automated operation with a pause at the beginning of the cycle. After the operator has been pushed the proceed button the automatic sequence will continue.

20

C. Cycle Editing:

When a chemist executes the cycle edit function from the Idle mode, he is first quiered for the method
25 number to edit. This method is then copied from the method buffers to the edit buffer. The editing screen/s are then presented to the chemist. The chemist can move between the edit fields to change the various method parameters. The cycle parameters are presented to the chemist and by using the NEXT and PREVIOUS keys changes the appropriate values in the appropriate units for the particular parameter being edited.

30 One function available during editing allows the chemist to save the changes back to the method buffer. When this function is executed, the entire method is validated, and then saved back to its method buffer. The editor is then exited, and the instrument is returned to the Idle condition.

A second function allows the chemist to print a method out. And a third function allows the chemist to modify the detector parameters.

35

1. Method Choices:

The following chart is a matrix showing the options and choices available when editing a run method.

40

45

50


55

5

Sample Method Possibilities						
Cycle	Time	Pause	Volt	Temp	Item	Other
Wash	X	X	N/A	N/A	N/A	N/A
Buffer	X	X	N/A	N/A	N/A	N/A
Marker	X	X	N/A	N/A	X	N/A
Sample	X	X	N/A	N/A	X	N/A
Detector	N/A	N/A	N/A	N/A	N/A	Rise, range, wave
T Zero	0.0	N/A	X	X	BUF	Polarity
T 1	X	N/A	X	X	BUF	Optional
T 2	X	N/A	X	X	BUF	Optional
T 3	X	N/A	X	X	BUF	Optional
T End	X	N/A				

10

15



Key: t=time T=Temperature

50

Notes:

Boxes 2-5 time decreases to zero for completion of task.

Boxes 7-11 time increase to chosen value for completion of task.

Box 6 MUST occur within 6 minutes with the temperature stable.

55

D. Method Memory:

The ACE instrument will have the ability to save four (4) methods in battery backed memory. This will allow the chemist to establish a set of methods and then maintain those methods even when the instrument is turned off. In addition to the four methods, there will be two buffers to hold the method currently being edited, and the method currently executing. This means there will a total of six (6) method buffers in the battery backed ram. A method buffer should consume approximately 30 bytes of memory.

10 E. Cycle Sequences:

The following section contains the sequence of events that are to take place for each operation. This is the start of the State definitions that the software will implement to create the desired use for the ACE unit.

15

i. Interlocks:

The following hardware interlocks prevent voltage from being applied to the capillary, rotation of the vial holder or up movement of the vial holder.

- 20 a. Buffer Door
- b. Detector Door
- c. Vial Holder Door
- d. Oven Access Door
- e. Vial Holder in Place
- 25 f. Vial Holder in Rotate allowed position

When the Doors can be opened or must be closed is indicated in the next two lists.

30 Doors May be Open

- 1. Idle Mode
- 2. Manual Mode
- 3. Pause

35

Doors Must be Closed

- 40 1. Voltage applied to capillary
- 2. Vial Holder Moving

i.a Interlock Sequence:

- 45 a. Look at Hardware Bit
- b. Look at Software Mirrors of processes allowed
 - 1. The software Mirrors are flags indicating the processes occurring at that time. They get changed at cycle transitions, cycle operation and exit from the cycle.

50

ii. Equilibrate Temperature:

- a. Read Oven Temperature from A/D and store value
- 55 b. Read Ambient Temperature and store value
- c. Compare Oven Temperature to setpoint

1. If Reading > setpoint
 - a. Delta Temperature > + 15 °C and ambient < Oven then open flap & turn off heater. Otherwise just turn heater off.

- 5 d. If Reading < setpoint then close flap & turn on heater.

Note: Checking ambient temperature and giving cautions for temperatures close to or lower than ambient will be a future design choice. Also in the future a valve to control coolant flow will be under software control.

10

iii. Vacuum Wash Sequence:

- 15 a. Move Vial Holder to Wash solution
- b. Make sure 5" and Inject valves are closed
- c. Open 20" valve
- d. Open Inject valve
- e. Wait X time [user selected time]
- f. Close Inject valve
- 20 g. Close 20" valve

iv. EquilibrateBuffer:

- 25 a. move vial holder to designated buffer number
- b. Vacuum buffer
- c. Wait X time [user selected time]

v. Marker Inject Sequence:

- a. move vial holder to marker solution
- b. perform user choice of vacuum or voltage
- c. Wait X time [user selected time]

35

v.b.1 Marker Vacuum Inject Sequence:

- 40 a. Make sure 20" and Inject valves are closed
- b. Check 5" vacuum level
- c. Open 5" valve
- d. Check 5" vacuum level
- e. Open Inject valve
- f. Wait X seconds [0 - 30 seconds user selectable]
- 45 g. Close Inject valve
- i. Close 5" valve
- j. Adjust 5" vacuum level

v.b.2 Marker Voltage Inject Sequence:

- a. Check Marker position
- b. Check Interlocks & in Up position
- c. Write to D/A voltage desired [Fixed value at this time]
- 55 d. Read A/D for current to indicate voltage conducting

vi. Sample Inject Sequence:

- a. move vial holder to sample number desired
- b. perform user choice of vacuum or voltage
- c. Wait X time [user selected time]

5

vi.b.1 Vacuum Sample Inject Sequence:

- a. Make sure 20" and Inject valves are closed
- b. Check 5" vacuum level
- 10 c. Open 5" valve
- d. Check 5" vacuum level
- e. Open Inject valve
- f. Wait X seconds [0 - 10 seconds user selectable]
- g. Close Inject valve
- 15 i. Close 5" valve
- j. Adjust 5" vacuum level

vi.b.2 Voltage Sample Inject Sequence:

20

- a. Check Marker position
- b. Check Interlocks & in Up position
- c. Write to D/A voltage desired [Fixed value at this time]
- d. Read A/D for current to indicate voltage conducting

25

vii. Detector Sequence:

- a. Set Wave Length
- 30 b. Set Range
- c. Set Rise Time

viii. t Zero Sequence:

35

- a. Interlocks OK
- b. At Temperature
- c. Turn on Integrator
- d. Turn on voltage [Stepped ONLY] and start run time
- 40 e. Send Autozero command to Detector

ix. t1 Sequence:

- 45 a. Go to next cycle if time is zero
- b. change voltage if non zero
- c. change temperature if non zero
- d. change buffer if non zero

50

ix. t2 Sequence:

- a. Go to next cycle if time is zero
- b. change voltage if non zero
- 55 c. change temperature if non zero

d. change buffer if non zero

x. t3 Sequence:

5

- a. Go to next cycle if time is zero
- b. change voltage if non zero
- c. change temperature if non zero
- d. change buffer if non zero

10

xi. t End Sequence:

- a. Turn off Voltage
- 15 b. Set temperature to what it was at time Zero
- c. turn off integrator
- d. Decrement number of samples to run
 - i. if number of sample = zero then go to Idle
 - ii. if number of samples not = zero begin autorun on next cycle. The autorun restarts at the
- 20 beginning of the method with temperature equalibration.

F. Specialized Sequences:

25

The sequences in this section are ones that are used by larger general sequences. Specialized sequences that can not find a home elsewhere are included here.

30 1. Capillary Vial holder Sequence:

The time from vial to vial is 1/4 of a second giving a worse case of 1 3/4 seconds to the next vial. The following chart shows the positions and coding for each of the eight possible vials.

35

Direction of rotation is ----->								
VIALS	Sam1	Sam2	Sam3	Sam4	Wash	Buf2	Buf1	Marker
Rough LED 1		x			x		x	x
Rough LED 2			x		x	x		x
40 Rough LED 3				x		x	x	x
Exact LED	x	x	x	x	x	x	x	x

45

1.2 Rotate Sequence:

- a. Sampler Down and at IN position
- b. Turn Motor On
- 50 c. Watch for Rough LED Pattern
- d. When Exact LED is ON turn Motor OFF
- e. Move Sampler Up

55 1.3 Up Vial Holder Sequence:

- a. check interlocks
- b. make sure IN position, not rotating & volts off

- c. move up
- d. stop when up

5 1.4 Down Vial Holder Sequence:

- a. Volts off, Vacuum off & interlocks OK
- b. Move Down
- c. Stop when indicator shows it is down

10

2. Power On Sequence:

- a. Initialize peripheral components
- b. Initialize ports
- c. Initialize vacuum
- d. Idle position vial holder
- e. Set Temperature to 30° C
- f. display Idle mode screen

15

20

3. Self Test Sequence:

- a. Choice run A test or Auto run tests
- b. Run test(s)
- c. Report result(s)
- d. Wait here for operator to go back to Idle menu

25

30 4. Power Fail Sequence:

- a. Save time occurred
- b. Save system status if needed
- c. Turn off voltage, vacuum, motor movement
- d. Wait for power to be restored
- e. query operator for choices of recovery

35

5. Real Time Clock Sequence:

- a. get command to be passed to chip
- b. send command
- c. get response if needed

40

45

6. Printer Sequence:

- a. Get pointer of data to be printed
- b. Initialize printer
- c. Get format for this data to be printed
- d. Start print data in proper format
- e. Complete printer operation

50

55 7. Buffer Bottle Status Sequence:

- a. Increment Buffer 1 & 2 number of times used counters
- b. If near depletion than give gentle caution

- c. If at or past depletion give firm warning
- d. reinitialize values as appropriate

5 8. Adjust 5" Vacuum Sequence:

- a. Read Vacuum A/D
- b. Compare to setpoint
- c. if reading > setpoint - XX then open bleed down valve.
- 10 d. if reading < setpoint + XX then open bleed up valve.

9. Power Down Sequence:

- 15 a. query operator for Choices
- b. save status according to operator choices
- c. put vial holder in store position.
- d. Give operator OK to power down.

20

10. Interrupt Service Routine:

- a. Check Hardware Interlock bit
- b. Increment Timer Counters
- 25 c. Future read ports to see they are at state desired

11. Round Robin Tasks:

- 30 a. Software Interlock Mirror Check (Future when can read ports)
- b. Voltage task
 - 1. Read A/D and convert to system value
 - 2. Adjust Voltage as needed
 - 3.
- 35 c. Vacuum task
- d. Temperature task
- e. Absorbance task
- f. Status display update
- 40 g. Chemistry
- h. Run Setup
- i. Edit
- j. Print
- k. RS-232
- 45 l. Manual Operation

III. Manual Control:

50 Manual control allows the chemist to perform any of the cycles of an automated run, but with complete manual control of the process. This will allow the chemist to test new ideas, and conditions in a manner where instant feed back to the changes are desired.

Manual Control will also have access to sub-state operations that are normally buried within an automated cycle. An example of such sub-cycles are the vial holder movement operations.

55 Manual control can be entered from two distinct places. The first place is the Idle mode. There will be a function key defined as manual control. The second place is from the Run mode. Again there will be a function key (preferable the same key) designated as Manual control.

When manual control is entered from the Idle mode, the instrument sets itself to the wash state of an

automated run, but does not begin execution of the cycle. It waits for the chemist to specify which cycle it should execute, and what the conditions are.

When manual control is entered from the Run mode, the instrument continues running the current cycle. If the tEnd cycle has been entered, the voltage, and temperature are frozen, and the instrument
 5 continues to operate at those conditions, until either the conditions are changed or stopped by the chemist. The chemist may choose to abort or allow the current cycle to complete.

When in manual mode, the chemist may select a cycle to execute. The chemist then enters values for that cycle's parameters. If all conditions for that cycle are met then it is executed. This includes, safety interlocks, positioning of capillary, and any other safety related items.

10

IV. Configuration:

Configuration allows the chemist to do two things. They can set the default values of instrument
 15 parameters and self test the instrument.

Default instrument parameters such as the detector parameters that do not change as often as other instrument values can be set here, and any time a default value for that parameter is required, the software gets from the system configuration.

The instrument self test allows the operator to force the instrument to perform its power up diagnostics.
 20 These will be defined as they are deemed necessary.

V. System Screens:

25

A. Idle Screen:

30	Line 1	Time TOD	Abs	Temp	Edit
	Line 2	Volts	Current	Y	Run
	Line 3				
	Line 4				
	Forty gage	1234567890123456789012345678901234567890			
		1	2	3	4

35

B. Edit with Run in process Screen:

40

40	Line 1	Method "Name"	Cycle "Name"	Main-
	Line 2	Temp XX	Volts XX	Run--
	Line 3-	Time XX	Buffer X	Print
	Line 4	Temp: XX	Volts XX	End XX
45	Forty gage	1234567890123456789012345678901234567890		Status
		1	2	3
				4

C. Manual Screen:

50

55

Line 1
Line 2
Line 3
Line 4
5 Forty gage 1234567890123456789012345678901234567890
1 2 3 4

10 VI. Error Handling:

There will be visual and audible error response to all errors. Each cycle that has error checking will have a section for error processing which will include message display and corrective action if possible and if not an appropriate shutdown procedure.

A. Error Screens:

20	Line 1	Interlocks Faulty Check Doors & Sampler																																							
	Line 2																																								
	Line 3																																								
	Line 4																																								
	Forty gage	1234567890	1234567890	1234567890	1234567890	1234567890	1234567890																																		
25		1	2	3	4																																				

VII. Hardware/Software Interface:

The information in this section will specify the relevant aspects of the hardware that the software needs for interface to the hardware. This includes port addresses of devices, control values of devices, and conversions for DACs, and ADCs.

A. Autosampler Controller:

Port: 73h

Autosampler consists of a pair of DC motors that are used to position a device. These motors are turned on and off, and the position is noted by monitoring a set of sensors. The bit pattern of the sensors indicates the position of the autosampler.

45 **B. Detector Controller:**

Port: 72h

At this stage the Detector Controller consists of an ADC to convert the absorbance signal from the back of the 783 detector to a digital value, and then display it. Other control of the detector will be via the Z80's RS = 232c port.

C. Temperature Controller:

Port: 71h

55 The temperature is controlled by reading the temperature value, comparing it to the setpoint value, and then turning the heater on/off and the opening/closing the flap to adjust the temperature up or down.

The following is the derivation of the temperature ADC reading to a temperature in °C.

5

10

15

0 to 5 V _{in}	=	0 to 60 °C
255 bits	=	60 °C
1 bit	=	0.253 °C
(ADC value * 253) / 1000 = T °C		
0 to 5 V _{in}	=	10 ° to 60 °C
255 bits	=	50 °C
1 bit	=	0.196 °C
(ADC value * 196) / 1000 = T °C		
0 to 5 V _{in}	=	15 ° to 60 °C per parameter temperature spec.
255 bits	=	45 °C
1 bit	=	0.1765 °C
(ADC value * 176.5) / 1000 = T °C		

The following is a list of valid values that can be written out to port 71h for controlling the heater.

20

00h	Turn heater off.
01h	Turn heater on.

25

D. Vacuum Controller:

30

Port:	74h	Injector Valves
	75h	5 inch Ballast Level

35

The vacuum level in the ballast tanks is read through port 75h. The level of the ballast is maintained by reading the vacuum level, and then bleeding the tank up or down to adjust it. The bleed valves are controlled by writing a bit pattern to port 74h. Each bit position in the byte represents one of the valves.

The following derivation converts the A/D reading of the ballast level to vacuum in psi.

40

0 to 5 V _{in}	=	0 to 15 inches
255 bits	=	15 inches
1 bit	=	.1197 inches of mercury
(ADC value * 119.7) / 1000 = inches		

45

The following is a list of valid values that can be written out to the value port (74h).

50

FBh	Turn On bleed-up valve.
F7h	Turn On Bleed-down valve.
EEh	Hi-Vacuum injector.
EDh	Lo-Vacuum injector.

55

E. Voltage Controller:

Port	70h	High Voltage Control
	77h	High Voltage Current

5 The high voltage power supply uses a 0 to 10 V input to produce a 0 to 30,000 V output. The DAC value for the output voltage is written to port 70h. The current across the capillary is read at port 77h.

The following is derives the equation to convert a voltage setpoint into an DAC value:

$$\begin{aligned}
 0 \text{ to } 10 \text{ V}_{in} &= 0 \text{ to } 30,000 \text{ V}_{out} \\
 1 \text{ V}_{in} &= 3000 \text{ V}_{out} \\
 255 \text{ bits} &= 10 \text{ V}_{in} \\
 1 \text{ bit} &= 0.0392156 \text{ V}_{in}
 \end{aligned}$$

$$\frac{1 \text{ V}_{in}}{3000 \text{ V}_{out}} = \frac{0.0392156 \text{ V}_{in}}{X \text{ V}_{out}}$$

$$X \text{ V}_{out} = 117.65 \text{ V}_{out}$$

$$117.65 \text{ V}_{out} \text{ per } 1 \text{ bit}$$

$$\text{setpoint} / 117.65 \text{ V}_{out} = \text{DAC value}$$

NOTE: setpoint is in increments of 1000 V

25 This derivation converts the high voltage ADC reading to a value in Volts.

0 to 4.95 V _{in}	=	0 to 0.33 mA
255 bits	=	4.95 V _{in}
1 bit	=	0.0194117 V _{in}

The next derivation converts the current ADC reading to a current in μA .

35 F. Polarity Relay:

Port: 74h value of 20h

Engaged relay (on) gives a minus charge to the receive buffer side of the cathode. Disengaged relay
40 (off) gives a plus charge to the receive buffer side of the cathode.

G. Event Relay:

45 Port: 74h value of 80h

This is the port used to turn on the Integrator.

H. Heater Flap Relay:

50 Port: 74h value of 40h

This port is used to let heat out of the box if it goes 15 °C over the desired temperature.

55 VIII. State Notes

1. Interlocks:

2. Equalize Temperature:

3. Vacuum Wash:
 4. Equalize Buffer:
 5. Marker Inject General Operation:
 6. Sample Inject General Operation:
 5 7. Inject Vacuum:
 8. Inject Voltage:
 Use $\text{Mod}(n1, n2, n3 - \Delta, n4) \rightarrow n1 \cdot n2 / n3$ Round up or down using remainder : setpt \rightarrow Code (<setpt - code>)
 Now rounded d/a setpoint code is on top of stack :S fetch setpt \rightarrow code dup codeI port pl ; Send it to D/A
 10 A/D value (00-FF) * 13 (constant) = move decimal one place to left and get current reading in micro amps.
 [0-330 μ A]
 9. t Zero Event Processing:
 10. Integrator:
 11. Start Automatic Operation:
 12. End Automatic Operation:
 15 13. Start/End Manual Operation:
 14. Rotate Vials:
 Enter: Value of vial desired
 a. At down yes/no
 20 no 1. Alarm error exit
 yes b. Get value of vial position
 c. Turn on Motor
 d. Read rough LED's wait for match yes/no
 no 1. loop here till match
 25 a. get time & save first entry
 b. see if time > than 2 sec. yes/no
 no loop & look for match
 yes Alarm error exit
 yes e. Read Exact LED wait for match yes/no
 30 no 1. loop here till match
 a. get time & save first entry
 b. see if time > than 1/8 sec. yes/no
 no loop & look for match
 yes Alarm error exit
 35 yes f. Turn Motor off
 g. Clear stack of time data
 h. Continue next sequence
 Exit: At vial location desired
 Error Exit:
 40 a. buzzer on
 b. display error message [See display routine]
 c. go to Idle mode [See Abort Sequence]
 15. Down Vials:
 45 16. Up Vials:
 17. ISR:
 18. Round Robin Operation:
 a. Software Interlock Mirror Check (Future when can read ports)
 b. Voltage A/D
 50 c. Vacuum A/D
 d. Temperature A/D
 e. Absorbance A/D
 f. Status display update
 g. Chemistry
 55 h. Run Setup
 i. Edit
 j. Print

k. RS-232

5

Claims

1. Apparatus for capillary electrophoresis comprising:
 - a first reservoir (21) for holding a first electrophoretic solution (19);
 - 10 a second reservoir (17) electrically isolated from the first reservoir for holding a second electrophoretic solution (15);
 - a sample reservoir (91) located in proximity of the first reservoir for holding a sample (93) to be electrophoresed;
 - a high voltage power supply (43) connected between the first reservoir (21) and the second reservoir (17);
 - 15 a first pressure source for providing an environment for the first reservoir and the sample reservoir of a first gas having a first known pressure, so that the first electrophoretic solution in the first reservoir and the sample in the sample reservoir are under the first pressure;
 - a pressure reservoir (99) for holding a second gas having a second pressure that is lower than the first pressure;
 - 20 a capillary tube (13) in which to electrophorese the sample;
 - holding means (57, 53) for holding the first reservoir and the second reservoir for holding the pressure reservoir, for holding the high voltage power supply, for holding the sample reservoir, and for holding one end of the capillary tube in the second reservoir;
 - connecting means (29) for connecting the second reservoir to the pressure reservoir, the connecting means
 - 25 comprising valve means (97) for venting the connecting means to the first pressure source and for blocking communication of the second reservoir with the pressure reservoir while venting to the first pressure source means and
 - insertion means (65) for inserting the other end of the capillary tube into the sample reservoir and into the first reservoir.
- 30 2. Apparatus according to claim 1, further comprising computer means (117) for controlling the insertion means and the valve means, so that when the other end of the capillary tube is in the sample reservoir, the valve means permits communication of the second reservoir with the pressure reservoir for a period of time for sucking the sample into the capillary tube.
3. Apparatus according to claim 2, wherein the computer means causes the other end of the capillary
- 35 tube to be transferred to the first reservoir after the sucking of the sample into the capillary tube.
4. Apparatus according to claim 3, wherein the computer means includes means for controlling the high voltage power supply to cause electrophoresis in the capillary tube after the other end of the capillary tube is transferred into the first reservoir.
5. Apparatus according to any one of claims 1 to 4, further comprising temperature control means (107)
- 40 for controlling the temperature of the capillary tube.
6. Apparatus according to claim 5, wherein the temperature control means is coupled to the computer means, and the computer means further comprises means for controlling the temperature control means.
7. A method of performing capillary electrophoresis comprising the steps of:
 - pre-washing a capillary;
 - 45 equilibrating the capillary with an electrophoretic medium to adjust the pH in the capillary until the charge on the inner wall of the capillary is substantially zero;
 - introducing a sample into the capillary;
 - applying an electric field to the capillary to cause electrophoresis; and
 - detecting the sample introduced into the capillary.
- 50 8. A method according to claim 7, wherein the step of prewashing the capillary is carried out using an acid.
9. A method according to claim 7 or claim 8, wherein the equilibrating step, the pH is adjusted to be less than 4.0.
10. A method according to claim 9, wherein in the equilibrating step, the pH is adjusted to be less than
- 55 2.5.
11. A method of performing capillary electrophoresis comprising the steps of:
 - introducing an electrophoretic medium into a capillary;
 - introducing a sample into the capillary;

applying an electric field to the capillary to cause electrophoresis;
adjusting the temperature in the capillary to vary the pH to cause differential separations in the sample during electrophoresis; and
detecting the sample introduced into the capillary.

5 12. A method according to claim 11, wherein the adjusting step comprises running the electrophoresis for a first period of time at a first temperature and then running the electrophoresis for a second period of time at a second temperature.

13. A method according to claim 11, wherein the adjusting step comprises changing the temperature in the capillary during electrophoresis according to a preselected temperature profile.

10 14. A method of introducing a sample into a capillary comprising;
placing one end of the capillary into a sample reservoir containing the sample, the sample in the sample reservoir being at a first pressure, the other end of the capillary being located in a second reservoir which is also at the first pressure; and

15 connecting the second reservoir to a pressure reservoir containing a gas having a second pressure lower than said first pressure for a controlled period of time to suck a portion of the sample into the capillary.

20

25

30


35

40

45

50

55



Run Number	Buffer	pH Level	Elution time Electroosmotic Flow (min)	Pre-Wash	Electroosmotic Mobility $10^{-4} \text{ cm}^2/\text{Vs}$
#1	Caps (10 mM)	pH = 11.13 pH = 11.13	t = 1.70 t = 1.63	NaOH HCl	8.2 8.5
#2	Bicene (10 mM)	pH = 8.45 pH = 8.45	t = 1.64 t = 1.93	NaOH HCl	8.5 7.2
#3	MES (10 mM)	pH = 6.0 pH = 6.0	t = 2.09 t = 6.27	NaOH HCl	6.7 2.2
#4	MES (10 mM)	pH = 6.0 pH = 6.0	t = 2.07 t = 4.80	NaOH HCl	6.7 2.9
#5	Citric Acid (20 mM)	pH = 4.0 pH = 4.0	t = 11.82 t = 54.6	NaOH HCl	1.2 0.25
#6	Citric Acid (20 mM)	pH = 2.51 pH = 2.51	t = 36.53 t = 70.31	NaOH HCl	0.38 0.2

Fig. 2





Fig. 3



Run Number	Buffer	pH Level	Elution time Electroosmotic Flow (min)	Pre-Wash	Electroosmotic Mobility $10^{-4} \text{ cm}^2/\text{Vs}$
#1	Caps (10 mM)	pH = 11.13 pH = 11.13	t = 1.70 t = 1.63	NaOH HCl	8.2 8.5
#2	Bicene (10 mM)	pH = 8.45 pH = 8.45	t = 1.64 t = 1.93	NaOH HCl	8.5 7.2
#3	MES (10 mM)	pH = 6.0 pH = 6.0	t = 2.09 t = 6.27	NaOH HCl	6.7 2.2
#4	MES (10 mM)	pH = 6.0 pH = 6.0	t = 2.07 t = 4.80	NaOH HCl	6.7 2.9
#5	Citric Acid (20 mM)	pH = 4.0 pH = 4.0	t = 11.82 t = 54.6	NaOH HCl	1.2 0.25
#6	Citric Acid (20 mM)	pH = 2.51 pH = 2.51	t = 36.53 t = 70.31	NaOH HCl	0.38 0.2

Fig. 2





Fig. 3



Run Number	Buffer	pH Level	Elution time Electroosmotic Flow (min)	Pre-Wash	Electroosmotic Mobility $10^{-4} \text{ cm}^2/\text{Vs}$
#1	Caps (10 nM)	pH = 11.13 pH = 11.13	t = 1.70 t = 1.63	NaOH HCl	8.2 8.5
#2	Bicene (10 nM)	pH = 8.45 pH = 8.45	t = 1.64 t = 1.93	NaOH HCl	8.5 7.2
#3	MES (10 nM)	pH = 6.0 pH = 6.0	t = 2.09 t = 6.27	NaOH HCl	6.7 2.2
#4	MES (10 nM)	pH = 6.0 pH = 6.0	t = 2.07 t = 4.80	NaOH HCl	6.7 2.9
#5	Citric Acid (20 nM)	pH = 4.0 pH = 4.0	t = 11.82 t = 54.6	NaOH HCl	1.2 0.25
#6	Citric Acid (20 nM)	pH = 2.51 pH = 2.51	t = 36.53 t = 70.31	NaOH HCl	0.38 0.2

Fig. 2

